

EU-DECLARATION OF CONFORMITY

Manufacturer	Wernli AG Eggasse 4 4852 Rothrist Switzerland
Single Registration Number CH:	CHRN-MF-20000003
EC Representative:	Wero Swiss Med Kft. Ipartelep utca 6 4220 Hajdúböszörmény Hungary
Single Registration Number HU:	HU-AR-000049642

We declare under our sole responsibility that

the medical device
with reference no
with Basic-UDI-DI (product code)

Wero Swiss Lan
see page 2
76300161A26PC

Risk Class

class 1, unsterile
According to Annex VIII, Rule 1

meets all the provisions of the Regulation (EU) 2017 /745 which apply to it.

Applied harmonised standards, national
standards or other normative documents

EN ISO 10993-1:2021 and applicable parts of the
EN ISO 10993 Series
EN ISO 15223-1:2021
EN ISO 14971:2019 / A11:2021

Conformity assessment procedures

Regulation (EU) 2017 /745
according to MDR Article 10 and MDR Article 52
(7) Annex I and Annex II

Place, date

Rothrist, 09.03.2026

Name and function



.....
Felix Schoenle, CEO

Reference Numbers:

Length: 5m stretched

A2620040.050	Wero Swiss Lan 4cm x 5m brown
A2620060.050	Wero Swiss Lan 6cm x 5m brown
A2620080.050	Wero Swiss Lan 8cm x 5m brown
A2620100.050	Wero Swiss Lan 10cm x 5m brown
A2620120.050	Wero Swiss Lan 12cm x 5m brown